



CERTIFICATE



SIRIM QAS International Sdn. Bhd. hereby certifies that

KARANGKRAF MEDICARE SDN. BHD.
LOT 1, JALAN RENGAM 15/5
OFF PERSIARAN SELANGOR
SEKSYEN 15
40200 SHAH ALAM
SELANGOR DARUL EHSAN
MALAYSIA



has implemented a Quality Management System for Medical Devices complying with

ISO 13485 : 2016

Medical Devices - Quality Management System – Requirements For Regulatory Purposes (In Compliance with Act 737 and Medical Device Regulations 2012 requirements)



Scope of Certification

MANUFACTURING AND DISTRIBUTION OF MEDICAL FACE MASK.

RISK CLASSIFICATION OF MEDICAL DEVICE: CLASS A, RULE 4

Issue Date : **07 March 2021**
Original certification date : **07 March 2021**
Expiry Date : **06 March 2024**
Certification No. : **QMS-MD 00179**

SIRIM QAS INTERNATIONAL SDN. BHD.
(Co. No. 410334 - x)
1, Persiaran Dato' Menteri
Section 2, P. O. Box 7035
40700 Shah Alam
Selangor Darul Ehsan
MALAYSIA
Tel : 06-3-55446404
Fax : 06-3-55446787

NUR FADHILAH BINTI MUHAMMAD
Chief Executive Officer
SIRIM QAS International Sdn. Bhd.

<http://www.sirim-qas.com.my>
<http://www.malaysiancertified.com.my>

This certificate is granted subject to the terms and conditions as stated in the Certification Agreement.

This Test Report refers only to samples submitted by the applicant to SIRIM QAS International Sdn. Bhd. and tested by SIRIM QAS International Sdn. Bhd. This Test Report shall not be reproduced, except in full and shall not be used for any purpose by any means or forms (including but not limited to advertising purposes) without written approval from the Head of Quality, Occupational Safety and Health & Environment (QOSHE), SIRIM QAS International Sdn. Bhd. Please refer to the last page of this Test Report for Conditions Relating to the Use of Test Report.

Test Results:

Product : 3 Ply Surgical Face Mask
 Brand : Karangkrat
 Model : IIR
 Size : 175mm x 95mm

No.	Type of Tests	Requirements BS EN 14683:2019 Clause 5.2 and Table 1: Performance Requirements for Medical Face Masks	Results	Remarks
1.	Bacterial Filtration Efficiency* (BFE), % (BS EN 14683:2019 Clause 5.2.2 and Annex B) <i>No of tested specimens: 5</i>	Type I : ≥ 95 Type II : ≥ 98 Type IIR : ≥ 98	> 99.99	Pass
2.	Differential Pressure (Breathability), Pa/cm ² (BS EN 14683:2019 Clause 5.2.3 and Annex C) <i>No of tested specimens: 5</i>	Type I : < 40 Type II : < 40 Type IIR : < 60	Average: 57.5	Pass
3.	Splash Resistance Pressure, kPa (BS EN 14683:2019 Clause 5.2.4 and ISO 22609: 2004) <i>No of tested specimens: 32</i>	Type I : Not required Type II : Not required Type IIR : ≥ 16.0	> 16.0 kPa (There was no sign of penetration, evidence of wetness or both appears on the viewing side of the total 32 specimens at 16.0kPa)	Pass
4.	Microbial Cleanliness*, cfu/g (BS EN 14683:2019 Clause 5.2.5) <i>No of tested specimens: 5</i>	Type I : ≤ 30 Type II : ≤ 30 Type IIR : ≤ 30	27	Pass

Note:

* Bacterial Filtration Efficiency (BFE) and Microbial Cleanliness test was conducted by SIRIM Industrial Biotechnology Research Centre (IBRC)



This Test Report refers only to samples submitted by the applicant to SIRIM QAS International Sdn. Bhd. and tested by SIRIM QAS International Sdn. Bhd. This Test Report shall not be reproduced, except in full and shall not be used for any purpose by any means or forms (including but not limited to advertising purposes) without written approval from the Head of Quality, Occupational Safety and Health & Environment (QOSHE), SIRIM QAS International Sdn. Bhd. Please refer to the last page of this Test Report for Conditions Relating to the Use of Test Report.

Test Results:

Product : 3 Ply Surgical Face Mask
 Brand : MediKraf
 Model : IIR
 Size : 175mm x 95mm

No.	Type of Tests	Requirements BS EN 14683:2019 Clause 5.2 and Table 1: Performance Requirements for Medical Face Masks	Results	Remarks
1.	Bacterial Filtration Efficiency* (BFE), % (BS EN 14683:2019 Clause 5.2.2 and Annex B) <i>No of tested specimens: 5</i>	Type I : ≥ 95 Type II : ≥ 98 Type IIR : ≥ 98	99.92	Pass
2.	Differential Pressure (Breathability), Pa/cm ² (BS EN 14683:2019 Clause 5.2.3 and Annex C) <i>No of tested specimens: 5</i>	Type I : < 40 Type II : < 40 Type IIR : < 60	Average: 56.4	Pass
3.	Splash Resistance Pressure, kPa (BS EN 14683:2019 Clause 5.2.4 and ISO 22609: 2004) <i>No of tested specimens: 32</i>	Type I : Not required Type II : Not required Type IIR : ≥ 16.0	> 16.0 kPa (There was no sign of penetration, evidence of wetness or both appears on the viewing side of the total 32 specimens at 16.0kPa)	Pass
4.	Microbial Cleanliness*, cfu/g (BS EN 14683:2019 Clause 5.2.5) <i>No of tested specimens: 5</i>	Type I : ≤ 30 Type II : ≤ 30 Type IIR : ≤ 30	3	Pass

Note:

* Bacterial Filtration Efficiency (BFE) and Microbial Cleanliness test was conducted by SIRIM Industrial Biotechnology Research Centre (IBRC)





INDUSTRIAL BIOTECHNOLOGY RESEARCH CENTRE
Building 19, SIRIM Complex

1, Persiaran Dato' Menteri, Section 2, P. O. Box 7035
40700 Shah Alam, Selangor Darul Ehsan, MALAYSIA
Tel: 603 - 5544 6953 / 6960 Fax: 603 - 5544 6988
Website: www.sirim.my

TEST REPORT

REPORT NO: R0051/21/B19/02

PAGE: 1 of 7

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to test item submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

Applicant : Karang kraf Medicare Sdn Bhd,
Lot 1, Jalan Renggam 15/5,
Off Persiaran Selangor,
Seksyen 15,
40200 Shah Alam,
Selangor Darul Ehsan,
Malaysia.

Manufacturer /
Company : Same as above

Test Facility: : Industrial Biotechnology Research Centre (IBRC),
Building 19, SIRIM Berhad.

Sample Name/
Trade Name: : 3-Ply Surgical Face Mask Brand: Karang kraf

Reference Standard /
Method of Test : ISO 10993-5: 2009. Biological evaluation of medical devices.
Part 5: Tests for *in vitro* cytotoxicity. (LWI-238-02)

Description of
Sample : Received one sample in good condition for testing with the following
identification:
1. Brand : Karang kraf
2. Lot / Batch No: Not provided
3. Manufacturing Date : Not provided
4. Expiry Date : Not provided
5. Quantity received : 50 pieces
6. Size : 175 mm x 95 mm
7. Color : White
8. Storage : Room temperature

Date Received : 12 January 2020

Job No. : J0051/21

Issue Date : 01 FEB 2021





INDUSTRIAL BIOTECHNOLOGY RESEARCH CENTRE

Building 19, SIRIM Complex

1, Persiaran Dato' Menteri, Section 2, P. O. Box 7035

40700 Shah Alam, Selangor Darul Ehsan, MALAYSIA

Tel: 603 - 5544 6953 / 6960 Fax: 603 - 5544 6988

Website: www.sirim.my

TEST REPORT

REPORT NO: R0052/21/B19/57

PAGE: 1 of 14

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to test item submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

Job No. : J0052/21

Applicant: : Karang kraf Medicare Sdn Bhd,
Lot 1, Jalan Renggam 15/5, Off Persiaran Selangor,
Seksyen 15,
40200 Shah Alam,
Selangor.

Company : Same as above

Sample Name : 3-Ply Surgical Face Mask - Brand: Karang kraf

Test Name : Primary Skin Irritation

Test Facility : Industrial Biotechnology Research Centre (IBRC),
Building 19, SIRIM Berhad.

Reference Standard : ISO 10993-10 (E)-3rd Edition 2010-08-01
(Exclude clause 6.5: Human Skin Irritation Test)
(LWI-238-57)

Receipt of Sample Date : 12 January 2021

Experimental Start Date : 17 January 2021

Experimental End Date : 22 January 2021

Issue Date : 25 January 2021



MS ISO/IEC 17025
TESTING
SAMM NO 576